



UA-2074 Ozone Testing Chamber

1. In compliance with ASTM D1171/ D1149, DIN/EN 27326, DIN 53509, JIS K-6259, ISO 1431/10960.
2. Ozone concentration accuracy: 2% of full scale.
3. Standard ozone range: 25~500pphm.
4. Temperature : -20°C~+100°C±1°C.
5. Includes Ozone detection and generation modules.
6. Functions : Dynamic and static tests.

10 - 11 Colrado Court

Office: Jl. Radin Inten II No. 62 Duren Sawit, Jakarta 13440 - Indonesia
Workshop: Jl. Pahlawan Revolusi No. 22B, Jakarta 13430 - Indonesia
Phone: 021-8690 6777 (Hunting)

Mobile: +62 816 1740 8925
Fax: 021-8690 6777